

****REPRESENTATIVE DATASHEET******Matched-Pair Antibody Set
for ELISA of human
Factor XIII antigen (FXIII)**Sufficient reagent for **5 x 96 well plates****Product #:** FXIII-EIA
Lot #: XXXX
Expiry Date: XXXX1395 Sandhill Drive, Ancaster, Ontario, Canada L9G 4V5
905-304-9896 • 800-903-6020 • fax 905-304-9897**Store at -10 to -20°C**For Research Use Only
Not for use in diagnostic procedures.**Description of Factor XIII (FXIII)**

Factor XIII (FXIII, fibrin stabilizing factor) is the proenzyme form of a transamidase that is essential for normal haemostasis and fibrinolysis, wound healing, female fertility and foetal development. Extracellular FXIII consists of A subunits (83 kDa each) which contain the enzyme moiety, and B subunits (76 kDa each) which act as a carrier protein for the A subunit in circulation. Both subunits are produced under separate genetic control. In plasma, FXIII exists as a non-covalent tetrameric complex (320 kDa) of two A-subunits and two B-subunits (A₂B₂). The concentration of FXIII tetramer in plasma is ~25 µg/ml (~80 nM). An intracellular form of FXIII is found in platelets, megakaryocytes and monocytes. This form of FXIII presents as a dimer of two A-subunits only and has a molecular weight of 160 kDa. The importance of these intracellular stores is demonstrated by the observation that platelets can contribute up to half of the FXIII activity in platelet rich plasma. The activation of FXIII involves several steps. Thrombin cleaves after Arg³⁷ of each A-subunit in the A₂B₂ tetramer, releasing a 4.5 kDa activation peptide. Additional conformational changes induced by the binding of calcium, and by dissociation of the B-subunits from the A-subunit dimer are required to obtain full enzyme activity. FXIIIa is a cysteine protease that catalyses the formation of γ-glutamyl-ε-lysyl bonds between the γ and α chains of polymerised fibrin molecules. Other proteins found crosslinked into fibrin clots by FXIIIa include fibrinogen, α₂antiplasmin, fibronectin, vitronectin and von Willebrand factor¹⁻³.

Principle of Sandwich-style ELISA

Affinity-purified polyclonal antibody to FXIII A subunit is coated onto the wells of a microtitre plate. Any remaining binding sites on the plastic wells are blocked with an excess of bovine serum albumin. The plates are washed and plasma or other fluids containing FXIII are applied. The coated antibody will capture the FXIII in the sample. After washing the plate to remove unbound conjugated antibody, the peroxidase activity is expressed by incubation with o-phenylenediamine (OPD). After a fixed development time the reaction is quenched with the addition of H₂SO₄ and the colour produced is quantified using a microplate reader. The colour generated is proportional to the concentration of FXIII in the sample.

Supplied Materials:

- 1. Capture Antibody (FXIII-EIA-C):** One yellow-capped vial containing 0.5 ml of affinity-purified polyclonal anti-FXIII A subunit antibody for coating plates.
- 2. Detecting Antibody (FXIII-EIA-D):** One red-capped vial containing 0.5 ml of peroxidase conjugated polyclonal anti-FXIII antibody for detection of captured FXIII.

Note: Antibodies are supplied in a 50% (v/v) glycerol solution for storage at -10 to -20°C. Keep vials tightly capped. Do not store in frost-free freezers.

Materials Required but not Provided:

This paired antibody set has been optimized for performance using the buffers and conditions described below. Most reagents are available as part of the VisuLize™ Buffer Pak (see Related Products section) or may be prepared as described below.

- 1. Coating Buffer:** 50 mM Carbonate
1.59g of Na₂CO₃ and 2.93g of NaHCO₃ up to 1 litre. Adjust pH to 9.6. Store at 2-8°C up to 1 month.
- 2. PBS:** (base for wash buffer and blocking buffer)
8.0g NaCl, 1.15g Na₂HPO₄, 0.2g KH₂PO₄ and 0.2g KCl, up to 1 litre. Adjust pH to 7.4, if necessary. Store up to 1 month at 2-8°C, discard if there is evidence of microbial growth.
- 3. Sample Diluent and Wash Buffer:** PBS-Tween (0.1%,v/v)
To 1 litre of PBS add 1.0 ml of Tween-20.
Check that the pH is 7.4. Store at 2-8°C up to 1 week.
- 4. Blocking Buffer:** PBS-BSA (1%, w/v)
Dissolve 2.5 g of Bovine Serum Albumin (Sigma-RIA grade) in 200 ml of PBS. Adjust pH to 7.4, if required, then make up to 250 ml with PBS. Aliquot and store frozen at -20°C.
- 5. Conjugate Diluent:** HBS-BSA-T20
5.95g HEPES (free acid), 1.46 g NaCl, 2.5 g Bovine Serum Albumin (Sigma, RIA grade) dissolved in 200 ml H₂O. Add 0.25 ml of Tween-20, check and adjust pH to 7.2 with NaOH, then make up to a final volume of 250 ml with H₂O.
Aliquot and store frozen at -20°C.
- 6. Substrate Buffer:** Citrate-Phosphate buffer pH 5.0
2.6g Citric acid and 6.9g Na₂HPO₄ up to a final volume of 500 ml with purified H₂O. Store at 2-8°C up to 1 month.
- 7. OPD Substrate:** (o-Phenylenediamine.2HCl) Toxic!
(5mg tablets: Sigma # P-6912). Make up immediately before use. Dissolve 5mg OPD in 12 ml substrate buffer then add 12 µl 30% H₂O₂. Do not store.

- 8. Stopping Solution:** 2.5 M H₂SO₄

Caution: VERY CORROSIVE! GENERATES HEAT ON DILUTION! Where stock sulphuric acid is 18 Molar, add 13.9 ml to 86 ml H₂O.
Store at room temperature.

9. Other:

Microplates, 96-well Immulon 4-HBX (<http://www.labsystems.fi>)
Microplate washer (optional)
Microplate reader.
Optional: VisuLize™ Buffer Pak
Optional: VisuCal™ Antigen Calibrator Plasma

Assay Procedure:

1. Coating of plates:

Dilute the capture antibody 1/100 in coating buffer (preferably in a polypropylene tube) and immediately add 100 µl to every well in the plate. Incubate for 2 hours at ambient temperature or overnight at 2-8°C.

2. Blocking:

Empty contents of plate and add 150 µl of blocking buffer to every well and incubate for 60 minutes @ 22°C.

Wash plate X 3 with wash buffer.

3. Samples:

Reference plasma is diluted 1/200 (100%) then serial 1/2's down to 1/6400 (3.13%). Sample plasmas are diluted 1/400, 1/800 & 1/1600. All dilutions are made in PBS-Tween sample diluent. Apply 100 µl/well and incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

4. Detecting Antibody:

Dilute the detecting antibody 1/100 in HBS-BSA-T20 conjugate diluent and apply 100 µl to each well. Incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

5. OPD Substrate:

Apply 100 µl of freshly prepared OPD substrate to every well. Allow colour to develop for 10-15 minutes then stop colour reaction with the addition of 50 µl/well of 2.5 M H₂SO₄. The plate can be read at a wavelength of 490 nm.

Calculation of Results:

The construction of a proper reference curve is of no less importance than any other aspect of the assay. A reference curve should be constructed by plotting the known concentration of standards versus absorbance. This can be done manually using graph paper, or by using curve-fitting computer software. In our experience, the dose response curves of most immunoassays tend to be sigmoid in shape. Although linear regions can be identified within the curve, the best overall fit is often obtained using an algorithm that provides a weighted theoretical model of fit throughout the entire curve, such as a 4-parameter or 5-parameter logistic curve fit^{4,5}. In general, the simplest model that defines the concentration-response relationship should be used⁶.

The "back-fit" test is a simple and reliable method to determine if a curve-fitting method is appropriate. In this test, the apparent concentrations for the absorbance values of each standard point are read from the reference curve. The derived values are compared to the assigned values. An appropriate curve fitting method will produce derived values that closely match assigned values throughout the range of the curve, within user-defined limits⁶. The coefficient of determination (R^2) is a valuable indicator of the overall fit, but should not be used by itself in the selection of a curve fitting method, as a poor fit in a particular region of the curve may not be evident from this value alone^{5,6}.

In the quality control of this product we have determined that under the conditions described above, a reference curve that is constructed using serial dilutions of normal pooled plasma, will produce a correlation coefficient (R^2) of at least 0.980 using a log-log fit, and an R^2 of at least 0.990 using a 4-parameter logistic curve fit algorithm. However, the performance characteristics of in-house assays developed using this product in other laboratories may vary slightly from ours. Different curve fitting methods may be employed but we recommend that the back-fit test be applied as evidence that the fitting method is appropriate.

Technical Notes:

- This paired antibody product is intended to facilitate the end user in establishing an in-house immunoassay for research purposes only. It must not be used for diagnostic applications. Assay validation is the responsibility of the end user and should be done according to user-defined protocols⁶.
- Reference calibrators should be of the same matrix and anticoagulant as the samples to be tested (example serum or plasma, citrate or EDTA).
- Do not use samples diluted less than 1/20, as falsely high readings may result.
- The optimal colour development time should be determined empirically as the time required to obtain an absorbance of at least 1.000 at 490 nm for the 100% reference point, not to exceed 20 minutes.
- Rheumatoid factor in samples may interfere in ELISA by binding to the capture and/or detecting antibodies.
- The wells should not be allowed to become dry. Keep plate covered or in a humid chamber during incubations.
- Antibodies are supplied in a 50% glycerol solution and can be centrifuged briefly in a micro-centrifuge to gather residual reagent from the cap and walls of the tube.

References:

1. McDonagh J; Structure and Function of Factor XIII; in Hemostasis and Thrombosis, 3rd Edition, eds. RW Colman, J Hirsh, VJ Marder and EW Salzman, pp 301-313, J.B. Lippincott Co., Philadelphia PA, USA, 1994.
2. Inbal A, Muszbek L; Coagulation Factor Deficiencies and Pregnancy Loss; Seminars in Thrombosis and Haemostasis 29, pp 171-174, 2003.
3. Murdock PJ, Owens DL, Chitolie A, Hutton RA, Lee CA; Development and Evaluation of ELISAs for Factor XIII A and XIII B Subunits in Plasma; Thrombosis Research 67, pp 73-79, 1992.
4. Nix, B, Wild D, in Immunoassays, A Practical Approach, editor J.P. Gosling, pp. 239-261, Oxford University Press, 2000.
5. NCCLS. Evaluation of the Linearity of Quantitative Analytical Methods; Proposed Guideline - Second Edition. NCCLS Document EP6-P2 (ISBN 1-56238-446-5, NCCLS, Wayne, Pennsylvania USA, 2001.
6. FDA Guidance for Industry. Bioanalytical Method Validation; May 2001, available on the internet: www.fda.gov/cder/guidance/index.htm

Related Products:

Cat #: EIA-PAK-1	VisuLize™ Buffer Pak: 5 plates, buffers, substrate
Cat #: EIA-CSA-1	VisuCal™ Antigen Calibrator, 1 x 1ml
Cat #: EIA-CSA-5	VisuCal™ Antigen Calibrator, 5 x 1ml
Cat #: SAFXIII-IG	Sheep anti-human FXIII, whole IgG from antiserum
Cat #: SAFXIII-HRP	Sheep anti-human FXIII, peroxidase labelled IgG
Cat #: SAF13A-IG	Sheep anti-FXIII-subunit A, whole IgG from antiserum
Cat #: SAF13A-AP	Sheep anti-FXIII-subunit A, affinity-purified IgG (APIgG)
Cat #: SAF13A-HRP	Sheep anti-FXIII-subunit A, peroxidase labelled IgG
Cat #: SAF13A-APFTC	Sheep anti-FXIII-subunit A, FITC conjugated APIgG
Cat #: SAF13A-APBIO	Sheep anti-FXIII-subunit A, biotinylated APIgG
Cat #: FXIII-DP	Human plasma deficient in FXIII, immune depleted
Cat #: FXIII-LDP	Human plasma deficient in FXIII, lyophilized

Visit our site (www.affinitybiologicals.com) for details.

Limited Warranty: This product is warranted to perform in accordance with its labelling and literature. Affinity Biologicals Inc. disclaims any implied warranty of merchantability or fitness for any other purposes, and in no event will Affinity Biologicals Inc. be liable for any consequential damages arising out of aforesaid express warranty.

Manufactured in Canada by:
AFFINITY BIOLOGICALS INC.
1395 Sandhill Drive
Ancaster, ON CANADA L9G 4V5
Tel: (905) 304-9896
(800) 903-6020
Fax: (905) 304-9897
info@affinitybiologicals.com
Document OPI0119, rev 3